

Visby Respiratory Health Test Information

Test Time	Approximately 30 minutes
Test Storage	Temperature: 36°F – 86°F (2°C – 30°C) Humidity: 5% – 80% DO NOT freeze
Operating Conditions	Temperature: 55°F – 88°F (13°C – 31°C) Humidity: 5% – 80%
Sample Type	Anterior swab (self or Health Care Provider collected) Nasopharyngeal swab, (HCP collected)
Patient Sample Storage	In dry tube up to 1 hour at room temperature 59°F–86°F (15°C–30°C) In buffer tube up to 2 hours at room temperature 59°F–86°F (15°C–30°C) In buffer tube up to 48 hours at refrigerated temperature 36°F–46°F (2°C–8°C) Do not freeze
Visby Respiratory Health Test 10 Pack	<ul style="list-style-type: none"> • 10 Visby Medical Respiratory Health Devices • 12 Pipettes • 12 Visby Respiratory Health Buffer Tubes • 12 Disposal Bags • 1 Quick Reference Guide
Materials Required and Available as Accessories	<ul style="list-style-type: none"> • Visby Power Adapter • External Controls
Materials Required but Not Supplied	<ul style="list-style-type: none"> • Gloves • Disposal Bin • Swabs
External Controls	Respiratory Health Control Swab Kit: <ul style="list-style-type: none"> • 2 Positive Swabs • 2 Negative Swabs
Need Help?	Email us: support@visby.com Call us: 1-833-468-4729 (1-833-GoVisby)