



Visby Medical Respiratory Health Test Ordering Checklist


This brief checklist is a resource that may be used to assist in ordering the Visby point-of-care Respiratory Health test.



Visby Medical
Respiratory Health Test



External Control
Swabs



Visby Power
Adaptor



Nasal Swabs

Ordering Supplies:

Visby Medical

- Tests** – Item # PS-400380 (20 devices per case)
- External Control Swabs*** – Item # PS-400381 (2 positive swabs, 2 negative swabs)
- Power Adaptor** – Item # PS-000288 (1 EA): Minimum 2 power adaptors are ideal in testing sites. Can be used with both the Visby Medical Respiratory and Sexual Health test. **TIP:** If a customer has high testing volume, more power adaptors may help workflow.

*QC is recommended to be run with new shipments and new untrained users and should only be purchased as needed.

Sold Separately

- Nasal Swabs** (anterior nasal or nasopharyngeal)
 - NOTE: Use sterile nylon, foam, or polyester flocked flexible shaft swabs, only. ≤80mm breakpoint preferred. Cotton and rayon swabs are not compatible.
 - Compatible with but not limited to:
 - Copan 502CS01 (AN) or 501CS01 (NP)
 - Puritan 25-3306-U (AN) or 25-3406-H (AN)

Training & Implementation Support

visby.training@visby.com



Customer Support

Hours: Mon-Fri
8:00 am – 4:00 pm PST

Email: support@visby.com

Phone: 1-833-GoVisby
(1-833-468-4729)